

# Outcome of proficiency test on EIA serology

#### Anses Dozulé Laboratory for equine diseases VIPARE unit

**Delphine Gaudaire** 


### EIA PT GENERAL INFORMATION

#### • AIM OF THE PT


Detect EIA antibodies:

using Agar Gel Immuno-Diffusion Test (AGID test) as described in the OIE

manual Chapter 2.5.6



using Enzyme ImmunoSorbent Assay (ELISA)





9/12/2018

### EIA PT GENERAL INFORMATION

- INFORMATION TO 34 LABS
  - Information sent by e-mail on May 17<sup>th</sup>, 2018
- SCHEDULE OF THE BOTH EIA PT
  - Samples sent on June 12<sup>th</sup>, 2018
  - Results due on July 6<sup>th</sup>, 2018
  - Each participant received an outcome of the PT by e-mail: EIA AGID PT: September 3<sup>rd</sup>, 2018
    EIA ELISA PT: September 4<sup>th</sup>, 2018
  - Official report will be sent to each lab by the end of September

### EIA PT CODIFICATION

• Lab codification:

	EIA AGID lab code	EIA ELISA lab code
Participant 1	18-A	18-G
Participant 2	18-B	/
Participant 3	18-C	18-D

• Codification of samples :

Lab	Sample #												
code	1	2	3	4	5	6	7	8	 20	21	22	23	24
<b>18-A</b>	645	267	288	426	95	656	632	691	 816	758	690	538	194
18-B	721	397	891	384	14	959	58	256	 572	533	794	191	958
18-C	407	569	893	172	623	776	355	405	 772	108	187	300	446

Anses – Dozulá Leboratory for eguine diseases 2018 EIA AGID PT

18-A-645

Reconstitute with 0,5 ml of distilled water

10th Workshops of equine diseases - Delphine GAUDAIRE

#### **EIA PT sample validation**


• Validation of each lyophilized serum:

Homogeneity

25 tested vials in 1 assay (FD ISO GUIDE 35) using ELISA

#### Reproducibility

3 vials tested in three independent AGID tests





9/12/2018

#### **EIA PT sample validation**

#### Stability

Short-term stability (stability after reconstitution):

- 3 days at room temperature
- 1 week at +2-8° C
- 1 month at -20° C

Long-term stability (accelerated conditions): 3 weeks at +30° C before reconstitution

## Homogeneity, reproducibility and stability have been validated for all the sera produced for the both PT

6


#### **Outcome of EIA AGID PT**



9/12/2018

### EIA AGID PT GENERAL INFORMATION


EIA AGID PT nb of participants: 32 (26 NRLs) •



EU COUNTRIES	EIA AGID PT 2010 2014 2018				
AUSTRIA					
BELGIUM					
BULGARIA					
CROATIA					
CZECH REPUBLIC					
CYPRUS					
DENMARK					
ESTONIA					
FINLAND					
FRANCE					
GERMANY					
GREECE					
HUNGARY					
IRELAND					
ITALY					
LATVIA					
LITHUANIA					
LUXEMBOURG					
MALTE					
POLAND					
PORTUGAL					
ROMANIA					
SLOVAKIA					
SLOVENIA					
SPAIN					
SWEDEN					
THE NETHERLANDS					
UNITED KINGDOM	04/05	0.4/05	00/05		
	21/26	24/28	26/28		

#### **EIA AGID PT samples**

Preparation of 9 batches of serum



9/12/2018

### EIA AGID PT samples

• 24 samples/panel to be analyzed by AGID test :

Serum #	Sample #	Expected result		
1	1 to 12	Negative		
2	13 to 14	Positive		
3	15 to 16	Positive		
4	17	Positive		
5	18	Positive		
6	19 to 20	Positive		
7	7 23 P			
8	24	Positive		
9	21 to 22	Positive		

4 strong positive: sample# 15-16-17-23

3 medium positive: sample# 18-21-22

5 weak positive: sample# 13-14-19-20-24


- ➤ 12 negative samples
- 12 positive samples

anse

- Criteria to evaluate the performance of each participant :
  - $\checkmark$  All negative samples should be found as such
  - ✓ All positive samples should be found as such
  - ✓ Only one mistake out of 24 samples was accepted

11


### EIA AGID PT results



• 100% of participants obtained satisfactory results

#### Each laboratory exhibited a very good specificity


### EIA AGID PT results




- 11 out of 372 positive samples have been analyzed as negative (specificity: 97%)
- 3 laboratories failed to detect positive samples

#### 29 out of 32 participants obtained satisfactory results

#### EIA AGID PT results





Unsatisfactory results are not in relation with the kit used

9/12/2018

10th Workshops of equine diseases - Delphine GAUDAIRE

anse

#### **EIA AGID PT conclusion**

- 29 out of the 32 labs obtained satisfactory results
- European NRL network exhibited a very good specificity 26/26 detected all negative samples
- The sensitivity of European NRL network is 88,5% 23/26 European NRL detected all positive samples
- 3 labs need to improve their sensitivity to detect all positive samples (especially weak positive sample)
- Training session will be organize by the EU-NRL


#### **Outcome of EIA ELISA PT**




9/12/2018

### EIA ELISA PT GENERAL INFORMATION

#### • EIA AGID PT nb of participants: 23




**EU COUNTRIES** AUSTRIA BELGIUM **BULGARIA** CROATIA CZECH REPUBLIC **CYPRUS** DENMARK **ESTONIA FINLAND** FRANCE GERMANY GREECE HUNGARY IRELAND ITALY LATVIA LITHUANIA LUXEMBOURG MALTE POLAND PORTUGAL ROMANIA **SLOVAKIA SLOVENIA** SPAIN SWEDEN THE NETHERLANDS UNITED KINGDOM



9/12/2018

#### EIA ELISA PT samples

Preparation of 9 batches of serum



9/12/2018

#### EIA ELISA PT samples

• 24 samples/panel to be analyzed by ELISA:

Serum #	Sample #	Expected result
1	1 to 8	Negative
2	9 to 11	Positive
3	12 to 14	Positive
4	15	Positive
5	16 to 18	Positive
7	19	Positive
6	20 to 21	Positive
8	22	Positive
9	23 to 24	Positive

- ➢ 8 negative samples
- 16 positive samples


anse

• Criteria to evaluate the performance of each participant :

- $\checkmark\,$  All negative samples should be analyzed as such
- ✓ All positive samples should be analyzed as such




### EIA ELISA PT results



100% of participants obtained satisfactory results

#### Each laboratory exhibited a very good specificity

### EIA ELISA PT results



 Z laboratories failed to detect positive sail (sensitivity: 98,4%)

#### 21 out of 23 labs obtained satisfactory results

#### EIA ELISA PT results



• Lab code 18-P and 18-S failed to detect the same serum (# 5)

#### 3 labs used INGENASA kits and 2 out of 3 failed to detect the positive serum 5



- 21 out of the 23 labs obtained satisfactory results
- European NRL network exhibited a very good specificity
- The sensitivity of European NRL network is 89,5% 17 out of 19 European NRL detected all positive samples
- 2 labs failed to detect all positive samples of the ELISA PT



### Acknowledgments



Aymeric HANS



#### Alexandre DESHIERE



Laurent HÉBERT







Jessy VANDEKERKHOVE

Fanny LECOUTURIER Anthony MADELINE



Investigate, evaluate, protect





10th Workshops of equine diseases - Delphine GAUDAIRE

anse

### Thank for your attention



9/12/2018